

Partial Discharge Tester

Sr. No.	Description	Qty
1.a	<p><u>PD Tester</u></p> <ul style="list-style-type: none">• Output voltage: 0.1 ~ 10 kV• Voltage accuracy: $\pm(1\% \text{ of setting} + 0.5\% \text{ full scale})$• Frequency: 50 Hz, 60 Hz• High voltage contact check function• PD measured result display• PD failure count setting• Power requirement: 220V, 50 Hz• Standard USB interface <p>Warranty: One year Brand: Megger, Chroma or equivalent Origin: USA/Japan/Europe</p>	01
1.b	<p><u>PD calibrator</u></p> <ul style="list-style-type: none">• Output voltage: 1 ~ 100 V• Polarity: negative• Power Supply: 9V battery	01

Note: Quote each item's price separately, where applicable. Clearly specify make, model numbers, and origin, above mentioned parameters and provide detailed literature in support. Failure to compliance may cause the rejection of the offer.

P:3

Check List

Partial Discharge Tester

Sr. No.	Description	Yes	No	N.A.
1	Part number of each item is mentioned			
2	Make, model no. and origin of each part is clearly mentioned			
3	Software is provided in the quotation			
4	Calibration certificate is provided for each parameter			
5	Calibration Kit is provided			
6	All standard accessories are mentioned			
7	Optional items prices are quoted separately			
8	Warranty period is clearly mentioned			
9	Inputs required to operate the items are mentioned			

Note: N.A = Not Applicable