

	TECHNICAL SPECIFICATIONS	TS-02
	Accessories for Condition Monitoring System & Eddy Current Testing Equipment	

S.No.	Technical Specification of Required Accessories	Quantity (Nos.)
1	Accelerometer (100 mv/g) with magnetic mounts compatible with CSI 2130 (Mil Type End)	08
2	Accelerometer cable 5 meter extension. (One-end BNC & other end Mil Type)	08
3	Accelerometer coiled cable 2 meter extension (One-end threaded & other Mil type)	01
4	02 Channel adapter (ACCEL A & ACCEL B) compatible with CSI 2130 Model 628A	02
5	04 Channel MUX Adapter compatible with CSI 2130 Model 648	02
6	25 pin female socket (built-in in CSI 2130)	02
7	CSI 2130 Batteries D24974 REV 5 (6) 1.2V NI-MH A CELL 7.2V/4.5AH	03

A) Compatibility	All the accessories (S.No. 1 to 7) should be compatible with software and hardware of main equipment (Model CSI 2130) of M/S Emerson Process Management
B) Service back up	Suppliers / manufacturer shall necessarily have service back up in Pakistan
C) Inspection	Inspection and testing at local supplier premises in Pakistan before delivery to site
D) Origin/Make	USA & Western Europe or equivalent.

Make: USA & Western Europe or Equivalent.

Equivalent Make / Model can be quoted too.

Technical literature / Operation & maintenance manuals (where applicable) in English language must also be provided.

All the accessories (Sr. No. 1 to 7) should be compatible with software and hardware of main equipment (Model CSI 2130) of M/s Emerson Process Management.