

Specifications of Differential Scanning Calorimeter

a) Differential Scanning Calorimeter with following specifications are required

Sr.#	Specifications	Requirement
1.	Maker	LINSEIS Messgeraete
2.	Model	Chip-DSC 10
3.	Quantity	01
4.	DSC Range	± 2.5 upto $\pm 250\text{mW}$
5.	Temperature Range	Room Temperature to 600°C
6.	Cooling System	*
7.	Heating Rate	*
8.	Cooling Rate	*
9.	Gas Flow Control	*
10.	Calibration Material	Included

* Manufacturer to specify

b) Scope of Supply

- i. The supplier will provide "Instruction/Operation Manuals & Video Tutorials" in English language.
- ii. Chinese or Taiwanese brands are not acceptable.
- iii. Any other prerequisite for use of this equipment.

Sr.No.	Description	Qty
1	2	3
1.	DIFFERENTIAL SCANNING CALORIMETER Maker: LINSEIS Messgeraete Model: Chip-DSC 10 DSC Range: ± 2.5 upto $\pm 250\text{mW}$ Temp. Range: Room temp. to 600°C Cooling System: * Heating Rate: * Cooling Rate: * Gas Flow Control: * Calibration Material: Included	01 No
2.	Grand Total (Rs)	

* Manufacturer to specify