Sr. #	Description	Unit	Qty.
	Semi-automatic Pull Tester Model Dage Series 4000 WP 100 OR Equivalent	No.	01
	X, Y and Z motorized axis		
	450mm X travel linear encoder stage		
	450mm Y travel linear encoder stage		
	X, Y Resolution < 1µm		
	Z Resolutions: ±0.1µm		
1	Test type: wire pull, shear and tweezers test		
1	Operating mode: manual and semi-automatic		İ
	Automatic report generator and data storage.		
-	Advanced Electronics and Software Control		
	Standard warranty of 12 months (Warranty will be started after installation and		Í
	commissioning of machine)		
	Accessories		
	1. Standards Accessories		
	Semiconductor multi-function cartridges P100 with compatible Hook (Wire		ł.
	pull range: from 0 to 100 g on wires from 18 to 75 µm)		
	3. Intel 8th Generation Processor based PC having Internal Windows® 10 with		
	Bond Test Manager, Mouse, Keyboard and 21 Inch color HD LED Monitor		
	4. Image Capture		
	5. Leica S6 microscope with 20x eyepieces, focusing arm, 6.3x to 40x		
	magnification		
	6. Installation and Training should be included	ļ	